

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS PO Box 1450 Alexasotra, Virginia 22313-1450 www.repto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/609,013	06/27/2003	Gregory R. Whittaker	022956-0216	8344
21125 7590 04/10/2008 NUTTER MCZENNEN & FISH I LP WORLD TRADE CENTER WEST 155 SEAPORT BOULEVARD BOSTON, MA 02210-2604			EXAMINER	
			ARAJ, MICHAEL J	
			ART UNIT	PAPER NUMBER
2001011,1111	DOGTON, INTO 2210 2004			•
			NOTIFICATION DATE	DELIVERY MODE
			04/10/2008	ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

docket@nutter.com

Application No. Applicant(s) 10/609.013 WHITTAKER ET AL. Office Action Summary Examiner Art Unit MICHAEL J. ARAJ 3733 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 20 December 2007. 2a) This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 1.3 and 5-19 is/are pending in the application. 4a) Of the above claim(s) _____ is/are withdrawn from consideration. 5) Claim(s) 11-19 is/are allowed. 6) Claim(s) 1.6 and 8-10 is/are rejected. 7) Claim(s) 3.5 and 7 is/are objected to. 8) Claim(s) _____ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. Attachment(s)

1) Notice of References Cited (PTO-892)

Paper No(s)/Mail Date _

Notice of Draftsperson's Patent Drawing Review (PTO-948)

31 Information Disclosure Statement(s) (PTO-1449 or PTO/SB/06)

Interview Summary (PTO-413)
 Paper No(s)/Mail Date.

6) Other:

Notice of Informal Patent Application (FTC 452).

Art Unit: 3733

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 1, 6 and 8-10 are rejected, as understood, under 35 U.S.C. 103(a) as being unpatentable over Taylor et al. (U.S. Patent No. 4,667,664) in view of

Taylor et al. disclose an adjustable drill guide comprised of a guide frame having an arm portion (42) and transverse base portion (36), a rod member (18) connected to the base portion and parallel to the arm portion further having an elongated stem portion and having an aperture (22) extending therethrough, and a guide member (46) including a channel (60) extending therethrough at an angle normal to the longitudinal axis of the arm as well as being moveable and lockable along the length of the arm portion. The lockable mechanism comprises a set screw (72). Taylor et al. also show the arm porting including indicia (78) representing the relative height of the channel with respect to the bone tunnel when the elongated stem portion is inserted inside the bone tunnel. With regard the statement of intended use and other functional statements, they do not impose any structural limitations on the claims distinguishable over Taylor et al. which is capable of being used as claimed if one so desires to do so. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law of anticipation does not require that the reference "teach" what the subject patent

Art Unit: 3733

teaches, but rather it is only necessary that the claims under attack "read on" something in the reference. Kalman v. Kimberly Clark Corp., 218 USPQ 781 (CCPA 1983).

Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. Ex parte Masham, 2 USPQ2d 1647 (1987).

Taylor et al. disclose the claimed invention except for the elongated stem portion having an enlarged head portion. Re et al. teaches the use of an enlarged head portion in order to facilitate tissue graft attachment (Page 3, paragraph 1). It would have been obvious to one skilled in the art at the time the invention was made to have created Taylor et al. with an enlarged head in view of Re et al., in order to have a device that would be able to be used in an easier manner.

Allowable Subject Matter

Claims 3, 5 and 7 objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Claims 11-19 allowed.

Response to Arguments

Applicant's arguments filed December 20, 2007 have been fully considered but they are not persuasive. Applicant argues that the examiner failed to provide the

Art Unit: 3733

required motivation for making the modification to the device disclosed in Taylor. In response to Applicant's argument that there is no suggestion to combine the references, the Examiner recognizes that references cannot be arbitrarily combined and that there must be some reason why one skilled in the art would be motivated to make the proposed combination of primary and secondary references. In re Nomiya, 184 USPQ 607 (CCPA 1975). However, there is no requirement that a motivation to make the modification be expressly articulated. The test for combining references is what the combination of disclosures taken as a whole would suggest to one of ordinary skill in the art. In re McLaughlin, 170 USPQ 209 (CCPA 1971). references are evaluated by what they suggest to one versed in the art, rather than by their specific disclosures. In re Bozek, 163 USPQ 545 (CCPA) 1969. In this case, it would have been obvious to combine Taylor et al. with an enlarged head in view of Re et al., in order to have a device that would facilitate tissue graft attachment.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the

Art Unit: 3733

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL J. ARAJ whose telephone number is (571)272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/Michael J Araj/ Examiner, Art Unit 3733

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733

Application Number